

Stanford University
W.W. Hansen Experimental Physics Laboratory
Gravity Probe B Relativity Mission
Stanford, California 94305-4085

GP-B Telescope Image Divider Assembly (IDA)
“Position Image Divider Housing
on Channel B Plate”
P0287 Rev -

June 30, 1997

Prepared: _____ Date _____
Ken Bower, Telescope Assembly

Prepared: _____ Date _____
Don Davidson, Telescope Assembly

Prepared: _____ Date _____
Jason Gwo, Telescope Assembly

Approved: _____ Date _____
Lynn Huff, Telescope Responsible Engineer

Approved: _____ Date _____
John Lipa, Telescope Manager

Approved: _____ Date _____
John Turneaure, Hardware Manager

Approved: _____ Date _____
Ben Taler, Quality Assurance

POSITION IMAGE DIVIDER HOUSING ON CHANNEL B PLATE

- for SUGP-B dwg# 25089
 - follows *Position Channel B Reflectors on Channel B Plate* (SUGP-B P0286)
 - also use *GP-B Telescope Image Divider Assembly (IDA) General Alignment and Bonding Procedures* (SUGP-B P0282).
- 1) Verify cleanliness of all fixturing.
 - 2) Mount fixture #506-0026A per OID dwg# 800-0047E.
 - 3) Position IDA assembly fixturing in workspace such that the end of the Channel B Roof Splitter can be monitored by autocollimation from a distant location (10'-20' away).
 - 4) Align remote autocollimator (Davidson Model 656) to end of Roof Splitter (<5 arc-seconds) and secure autocollimator and assembly fixturing as possible. Once bonding has begun, it will no longer be possible to verify correct alignment of beam. It may be necessary to shutdown all fans and otherwise minimize vibrational disturbances to achieve this alignment. Keep all air filtration systems operating whenever possible.
 - 5) Place a small piece (~3cmsq.) of 2mil 'orange' shim stock over each corner of Channel B Plate to protect surfaces during initial alignment.
 - 6) Verify cleanliness of all fixturing.
 - 7) Place Image Divider Housing (SUGP-B dwg# 25082) on shims on Channel B Plate assembly. Use care to prevent damage to flight parts.
 - 8) Coarsely adjust three brass screws such that the Housing is visibly aligned with the Channel B Plate assembly (SUGP-B dwg# 25444) (<0.020").
 - 9) Adjust the screw parallel to the knife edge such that the channel B 'face' of the Housing is closely aligned (<0.005") to the corresponding 'flat' on the Plate.
 - 10) Turn on the laser to maximum power. Place a small piece of translucent material (an ~1"x3" piece of clean room wipe works well) over the channel B 'face' such that the laser spots and outlines of the through holes in the Housing are visible (background illumination is helpful). Use care to not damage the flight part or contaminate a bonding surface.
 - 11) Adjust the two screws perpendicular to the Roof Prism edge such that the position of the two laser spots is centered (<0.010") with respect to the position of the two through holes. (Under current design, this yields a 'cross-eyed' appearance and each spot will not be centered with respect to its hole.)
 - 12) Adjust the screw farthest from the autocollimator such that the return image from the channel B 'face' is aligned (<10 arc-seconds) to the return image from the end of the Roof Splitter.
 - 13) Repeat steps 9-12 as necessary to final alignment.
 - 14) Remove Housing to verify autocollimator to Roof Splitter end alignment. If there is any change, repeat steps 4 and 13.
 - 15) Verify cleanliness of all fixturing.

- 16) Remove Housing and shims. Bond Housing to Channel B Plate using *Bonding Procedures for Fused-Quartz Components* (SUGP-B P0218).
- 17) Continue to monitor laser spots and autocollimator image after bonding and make alignment adjustments as required within a few minutes.

18) After curing, remove fixture #506-0026A. Remove completed Image Divider Housing/Channel B Assembly.

Attachments: SUGP-B dwg#'s 25089, 25082, 25444; OID dwg# 800-0047E.